



Droplet Measurement Technologies Calibration Data




**DROPLET
MEASUREMENT
TECHNOLOGIES**

US Patent
#7,656,510

MODEL NO.: CCN-100
SERIAL NO.: 2007-179
MFG. DATE: July 2020

Longmont, CO, USA
www.dropletmeasurement.com



**DROPLET
MEASUREMENT
TECHNOLOGIES**

Calibrated 7/2020
Calibration Due 7-2021
By DSJ

Longmont, CO, USA
www.dropletmeasurement.com



**DROPLET
MEASUREMENT
TECHNOLOGIES**

CERTIFICATE OF CONFORMANCE

Droplet Measurement Technologies, Inc.

certifies that the

**Cloud Condensation Nuclei Counter-Single Column
(CCN-100/AAA-0024)
S/N: 2007-179**


has received factory optical cleaning, optical alignment and calibrations. The instrument is certified to meet all factory calibration tolerances in accordance with DMT standards as of the date of this certificate.

Certificate Date.....7/2020


Sales Order.....1706050

Expiration Month7/2021

Calibration by:


D Johnson
CCN Product Engineer

Verified by:


C Puga
Service Engineer

Atmospheric Particle Spectrometers are used in very harsh environments and require routine cleaning and calibration verification during campaigns. The user is responsible for routine maintenance. DMT advises a full factory optical cleaning, optical alignment and calibration adjustment once per year.



2400 Trade Center Avenue
Longmont, CO USA 80305
Tel: 303-440-5576
www.dropletmeasurement.com
customer-contact@dropletmeasurement.com



Certificate of Calibration

Customer: LACy
Instrument: CCN-100
Serial Number: 2007-179
Invoice Number: 1706050

Field Office: Longmont, Colorado
Temperature: 25°C
Humidity: 42%RH

Instrument Status

Customer Recommended Calibration: Calibration Due: 7/2021
DMT Recommended Calibration: XX **Customer Invoice Number:** ADDAIR

Instrument Condition Received:

In Tolerance: Operational Failure: New: XX
 Out of Tolerance: Physical Damage:

Out of Tolerance Description:

Instrument Condition Shipped:

Meets DMT Specifications: XX Other (description attached):

Standards Used:

Particles


<u>Manufacturer</u>	<u>Particle Size</u>	<u>Standard Deviation</u>	<u>Lot No.</u>	<u>Expiration Date</u>
---------------------	----------------------	---------------------------	----------------	------------------------

<u>Flowmeter</u>		
<u>Type</u>	<u>Manufacturer</u>	<u>Serial Number</u>
BUCK mini Calibrator	AP Buck	A55022

<u>Reference Unit</u>		
<u>Type</u>	<u>Manufacturer</u>	<u>Serial Number</u>
SEMS	Brechtel	02
Thermometer	Fluke	00356720012S

Calibrated by: David Johnson **Calibration Date:** 7/20/2020

Quality checked by: Chuck Puga **Date:** 7/20/20

	Droplet Measurement Technologies Inc. 2400 Trade Centre Avenue, Longmont, CO, 80503, USA Ph: 303-440-5576, Fax: 303-440-1965 www.dropletmeasurement.com		CCN CAL Data
Customer: LACy	Project: New Instrument	Work Order No.: C179	
Serial No. CCN: 2007-179	Technician: dsj	Date: 7/20/2020	
Labels: <input checked="" type="checkbox"/>	Serial Tag: <input checked="" type="checkbox"/>	Software Rev Level: 6.0.2.100	
OPC SN: H311	Win 7 SN: 02219-088-190-640	CDPE Firmware Rev: 20120418BBCN.mcs	
Computer Speed: 1GHZ	Hard drive: 128gb		
Leak Rate: < 1 mb per sec.	Power Supplies: <input checked="" type="checkbox"/>	0Ω Between All GNDs: <input checked="" type="checkbox"/>	
Set 2.17V: <input checked="" type="checkbox"/>	Check +7V: 7.54V	Check Fuse Values: <input checked="" type="checkbox"/>	
Fans: <input checked="" type="checkbox"/>	LEDS: <input checked="" type="checkbox"/>	Diamond card Calibration: <input checked="" type="checkbox"/>	
Power Resistors: 3ohm 50W	Power switch: 20A		
D.C. Voltage Set: <input checked="" type="checkbox"/>	D.C. Voltage Read: <input checked="" type="checkbox"/>	Temperatures Calibrated: <input checked="" type="checkbox"/>	
Solenoid Pumps: 60ML <input checked="" type="checkbox"/> 20ML <input type="checkbox"/>			
Valve Min: 0.1	Valve max: 4.9		
Sample Flow Slope: 66.16	Sheath Flow Slope: 736.30	Pressure Slope: 221.45	
Sample Flow Int.: -163.69	Sheath Flow Int.: -1845.73	Pressure Int.: 109.84	
Current T1: 3.19A	Current T2: 3.19A	Current T3: 3.20A	
T1 Gain: 1.985	T2 Gain: 1.989	T3 Gain: 1.993	
T1 Offset: 0	T2 Offset: 0	T3 Offset: 0	
TG %: 7	DMA Calibration Slope: 18.86	DMA Calibration INT: .99	
P.I.D, Values: P:1 I:1 D:0			
Laser Power: 23mW	Laser Current: 87mA	Current limit: 3.53V	
Current Alarm: 105 mA	Baseline 0 level: 2.71V	Baseline calibrated: 2.81V	
OPC 1 st Stage: .04V	CDPE +-5V: 5.15V	1 st stage Rail: 5.00V	
Chiller ON: 22K	Chiller OFF: 26K	OPC Air Pump: <input checked="" type="checkbox"/>	
Water Pumps: <input checked="" type="checkbox"/>	60 Min Zero Count: <input checked="" type="checkbox"/>	48 Hour burn in: <input checked="" type="checkbox"/>	
Cold Turn on: <input checked="" type="checkbox"/>			
USB Port: <input checked="" type="checkbox"/>	Data Out Port: <input checked="" type="checkbox"/>	Ethernet Port: <input checked="" type="checkbox"/>	
Touch Screen Cal.: <input checked="" type="checkbox"/>	Computer Directory: <input checked="" type="checkbox"/>	Computer Start Menu: <input checked="" type="checkbox"/>	
Keyboard Verification: <input checked="" type="checkbox"/>	28V PS verification: <input checked="" type="checkbox"/>		
Bottle Caps Secured: <input checked="" type="checkbox"/>	Column Dried: <input checked="" type="checkbox"/>	Final Visual Inspection: <input checked="" type="checkbox"/>	
Wires Dressed neatly: <input checked="" type="checkbox"/>	USB Stick: <input checked="" type="checkbox"/>	Torque seal Pots: <input checked="" type="checkbox"/>	

Accessories:

KIT-0112 CCN-100 Consumables

KIT-0113 CCN-100 Basic Field Service

Comments:

System has 60 micro liter supply, drain, and OPC pumps.

Windows 7 based computer.

System has ABD-0234 Rev. C (CDPE DSP board).

100% tested as a system.

Warranty of repairs:


Warranty of 6 months on any repair work performed or parts replaced (except consumable items). Customer pays shipping to DMT. DMT will pay return shipping via standard ground.

CCN Flow Calibration Worksheet

CCN SN:	179	DATE:	7/20/2020	Tech:	dsj		
NOTE: take 5 points only in range of 20-75, one must be 45 +-2, RSQ must be >.99							
Sample flow							
				slope	66.16469		
				intercept	-163.694		
Volts	Flow			rsq	0.999495		
2.78	20.32					20	
2.97	32.66					32	
3.16	45.81					45	
3.4	60.51					60	
3.6	74.91					75	
Note: take only 5 points in range of 400mb to 1000mb, one will be around 840, one will be around 1000, other 3 will be less than 840, RSQ.must be >.99							
Pressure calibration							
				slope	221.4551		
				intercept	109.8387		
Volts	press mbar			rsq	0.999875		
1.3	400					400	
1.82	510					510	
2.3	620					620	
2.81	730					730	
3.32	847					840	
4	1000					1000	
NOTE: take 5 points only in range 200-750, one must be 450 +-10, RSQ.must be >.99							
Sheath flow							
				slope	736.30127		
				intercept	-1845.729		
Volts	sample	sheath	total	rsq	0.9913296		
2.73	16.45	183.55	200			200	
2.94	29.15	296.35	325.5			325	
3.07	37.67	412.83	450.5			450	
3.27	50.76	549.44	600.2			600	
3.41	67.25	682.95	750.2			750	
Set total flow to 500, flow ratio to 10 and measure flows, sample must be +-2.5%, Total must be +-5%							
			Flow Meter	CCN Display	%		
Sample			45	45	0.00		
Total			500	500	0.00		
Max Flow ratio Must be 10.6-10.8, adjust restriction if needed							
Sample flow rail >150							
5/6/2015 dsj							


10pt Temperature VS. SS% calibration flow 500vccm and 840mb

target delta T	T1	T2	T3	delta T: (T2-T1)	delta T: (T3- T1)	Kohler'Lookup T: 1/2*(T1+T2)	d50	SS%	delta T: 2*(T2- T1)
2	26.78	27.85	28.68	1.1	1.9	27.3	193.0	0.05251	2.1
4	27.38	29.40	31.13	2.0	3.8	28.4	97.0	0.14981	4.0
6	26.69	29.85	32.48	3.2	5.8	28.3	63.5	0.28912	6.3
8	28.10	32.12	35.53	4.0	7.4	30.1	51.5	0.39488	8.0
10	28.61	35.71	37.97	5.1	9.4	31.2	44.5	0.49177	10.2
12	28.51	34.56	39.79	6.0	11.3	31.5	39.0	0.60325	12.1
14	27.85	34.89	40.91	7.0	13.1	31.4	35.0	0.71562	14.1
16	28.23	36.42	43.37	8.2	15.1	32.3	32.5	0.79827	16.4
18	29.08	38.20	46.00	9.1	16.9	33.6	30.0	0.89606	18.2
20	28.49	38.62	47.32	10.1	18.8	33.6	27.5	1.02807	20.3




Final calibration data CCN-100
SN:2007-179 7/20/2020

Zero degree




Final calibration data CCN-100
SN:2007-179 7/20/2020

Ambient scan




Final calibration data CCN-100
SN:2007-179 7/20/2020

Final temperature calibration



Final calibration data CCN-100
SN:2007-179 7/20/2020

10 Degree gradient



Final calibration data CCN-100
SN:2007-179 7/20/2020

Calibration Coefficients

CCN Calibration Editor v1

Adjust the parameters below to the desired values for this instrument. Press Save to write the displayed parameters to the Current Configuration File or press Cancel to close this window without changing the Current Configuration File.

Save Cancel

Pressure Temp Coeff A

66.16 Sample Slope

-163.69 Sample Y-intercept

726.3 Sheath Slope

-1845.72 Sheath Y-intercept

18.86 Temp Gradient Slope

0.99 Temp Gradient Intercept

221.45 Abs Press Slope

109.84 Abs Press Y-intercept

Alarm threshold (low cts) 10

Monitor Alarm

4.9 Valve Set Max

0.1 Valve Set Min

2 T1-Sample

-1 Traction-T1

1 Traction-T3

2 TPOC-T3

1.1 offset T1 gain 1.985

0 T2 offset T2 gain 1.989

0 T3 offset T3 gain 1.993

7 %TG

Warning DMRT factory settings do not adjust without contacting DMRT

105 Max Laser Curt

Auto SS Scaleneil Card Installed

10 Last Delta T

2 Default SS

19.85 TG Dwn

Liquid Supply Pump 60 microliters

1.0000 P value

1.0000 I value

0.0000 O value

1.0000 Flow Avg #


DropLet MEASUREMENT TECHNOLOGIES

v1.6.0.0

Hide DMRT Settings

Final calibration data CCN-100
 SN:2007-179 7/20/2020

Zero Count at 500 mb



Final calibration data CCN-100
SN:2007-179 7/20/2020