

Picarro Certificate of Compliance

APPLICATION: CO2, CH4, CO and H2O
 DAS SN: 3664-CFKAD52372

PARAMETER	SPECIFICATION	VALUE	UNITS	Pass/Fail	NOTES
CO2 PRECISION - 5 MIN AVG	<= 50	17	ppb	Pass	Calculate 300 second interval standard deviation of raw data over 24 hrs. (1-sigma); with one hour stabilization at beginning of run
CO2 DRIFT	< 20	7	ppb	Pass	Calculate standard deviation of 5 minute averaged data over 2 hrs. (1-sigma); with one hour stabilization at beginning of run
CO2 ACCURACY	+/- 1.00	23	ppb	Pass	Calculate 50 minute interval mean of raw data over 30 hrs. (peak to peak); with one hour stabilization at beginning of run
CO2 REPRODUCIBILITY	<= 50	-0.08	ppm	Pass	Mean value of raw data over 30 minutes (average); with one hour stabilization at beginning of run (relative to Picarro standard) Cycle between two gases and ambient air at 20 minute intervals for 72 hrs. average 8 minutes 50 seconds to 9 minutes of each interval and calculate the standard deviation of averaged results for each gas. (1-sigma); with one hour stabilization at beginning of run
CH4 PRECISION - 5 MIN AVG	<= 1	0.15	ppb	Pass	Calculate 300 second interval standard deviation of raw data over 24 hrs. (1-sigma); with one hour stabilization at beginning of run
CH4 DRIFT	< 1.0	0.10	ppb	Pass	Calculate standard deviation of 5 minute averaged data over 2 hrs. (1-sigma); with one hour stabilization at beginning of run
CH4 ACCURACY	+/- 4.5	0.22	ppb	Pass	Calculate 50 minute interval mean of raw data over 30 hrs. (peak to peak); with one hour stabilization at beginning of run
CH4 REPRODUCIBILITY	< 0.5	-1.1	ppb	Pass	Mean value of raw data over 30 minutes (average); with one hour stabilization at beginning of run (relative to Picarro standard) Cycle between two gases and ambient air at 20 minute intervals for 72 hrs. average 8 minutes 50 seconds to 9 minutes of each interval and calculate the standard deviation of averaged results for each gas. (1-sigma); with one hour stabilization at beginning of run
CO PRECISION (5 minute average)	<= 15.0	2.8	ppb	Pass	Calculate 300 second interval standard deviation of raw data over 24 hrs. (1-sigma); with one hour stabilization at beginning of run
CO DRIFT	< 10.0	0.3	ppb	Pass	Calculate 50 minute interval standard deviation of 5 minute average data over 24 hrs. (1-sigma); with one hour stabilization at beginning of run
CO 72 HR DRIFT	<= 2	0.4	ppb	Pass	Peak to peak measurement variation with a five minute average over 24 hrs
CO ACCURACY	+/- 1.00	-0.04	%	Pass	2 x 1 Hr interval standard deviation collected with reference gas (5 minute reference/15 minute analysis) Mean value of raw data over 30 minutes (average); with one hour stabilization at beginning of run (relative to Picarro standard) Cycle between two gases and ambient air at 20 minute intervals for 72 hrs. average 8 minutes 50 seconds to 9 minutes of each interval and calculate the standard deviation of averaged results for each gas. (1-sigma); with one hour stabilization at beginning of run
CO REPRODUCIBILITY	< 1	0.5	ppb	Pass	Calculate 300 second interval standard deviation of raw data over 30 hrs. (1-sigma); with one hour stabilization at beginning of run
H2O PRECISION	< 30	8	ppm	Pass	Calculate standard deviation of 5 minute averaged data over 2 hrs. (1-sigma); with one hour stabilization at beginning of run
H2O PRECISION - 5 MIN AVG	< 5	1	ppm	Pass	Measurements for all three gas species made over this time interval and corresponding concentration levels assigned
MEASUREMENT INTERVAL	<= 5	2.4	seconds	Pass	10 to 90%; Configure the analyzer to measure CO2 only (1 second measurement interval typical); average of five transitions. Cycle between two gases at 5 minute intervals for 30 minutes; minimum delta between gas bottles is 20 ppm
RISE TIME	<= 5.0	2.4	seconds	Pass	90 to 10 %; Configure the analyzer to measure CO2 only (1 second measurement interval typical); average of five transitions; Cycle between two gases at 5 minute intervals for 30 minutes; minimum delta between gas bottles is 20 ppm
FALL TIME	<= 5.0	2.3	seconds	Pass	Absolute value of change in accuracy of gas species after 5-40C ambient temperature cycles 1.25 hr after restart of analyzer; measure accuracy for 15 minutes
GAS FLOW RATE	<= 280	229	scfm	Pass	Absolute change in accuracy of gas species after 5-40C ambient temperature cycles 1.25 hr after restart of analyzer; measure accuracy for 15 minutes
CO2 STORAGE TEST	<= 150	40	ppb	Pass	Absolute change in accuracy of both gas species after 5-40C ambient temperature cycles 1 hr after restart of analyzer; measure accuracy for 15 minutes
CH4 STORAGE TEST	<= 1.0	0.7	ppb	Pass	
CO STORAGE TEST	< x x	0.3	ppb	Pass	

GENERAL COMMENTS: * CO accuracy is measured relative to Picarro CO cylinder at 1000 ppm

Approvals
 Manufacturing
 Engineering
 Marketing

Initials
 SMD
 EIW
 GL

Date
 5/12/2020
 5/12/2020
 5/12/2020

CO2 CFKADS V1.0

